

AP-IgG Conjugation Kit For labeling 1 x 10 mg IgG

REAGENT STORAGE

The kit is shipped on blue ice. Please store kit components as described below.

| Kit Component | Storage Temp | Storage Notes |
|------------------------|----------------|--|
| Concentrated Activator | -20°C | Keep the vial in the desiccated container as supplied in the kit |
| AP-Z TM | -20°C or 2-8°C | Does not need to be kept desiccated. |
| Quenching Reagent | -20°C or 2-8°C | Does not need to be kept desiccated. |

INTRODUCTION

Alkaline Phosphatase is widely used as a fluorescent label in immunochemistry techniques such as flow cytometry and cellular analysis. Preparation of bright, stable, and reproducible antibody-AP conjugates is one of the biggest challenges for development of high-quality fluorescent reagents for cellular analysis and flow cytometry. The AP-IgG conjugation kit utilizes a novel chemistry to generate bright and highly reproducible AP-IgG conjugates with a simple procedure

FEATURES

- Liquid-based reagents.
- Completely scalable: conjugate anywhere from 10 μg to 1-gram IgG per reaction.
- Supplies sufficient activated AP to conjugate all IgG at a 2:1 AP:IgG ratio.
- Highly efficient AP incorporation purification not usually necessary.
- Customize the AP:IgG ratio to create optimized conjugates for different applications.
- Conjugates have greatly improved stability vs Lightning-LinkTM and traditional chemistry.

PRODUCT and CONTENTS

| Catalog Number | CNJ93103-010 |
|--|---------------|
| For Labeling: | 1 x 10 mg IgG |
| Concentrated Activator | 10 μL |
| AP-Z [™] - Activated AP (7.4 mg/ml) | 2 mL |
| 1X Quenching Reagent | 400 μL |

ADDITIONAL REAGENTS REQUIRED BUT NOT SUPPLIED

1X Phosphate Buffered Saline (1X PBS), pH 7.2-7.5 Deionized water (dH₂O) 1.5 ml microcentrifuge tubes



SHELF LIFE

The performance of the product is guaranteed for a minimum of 12 months when stored as directed.


IgG Requirements

The IgG to be labeled should be at a minimum concentration of 0.8 mg/ml in pure 1X PBS and should not contain any preservatives or carriers such as sodium azide, Proclin 300 or BSA.

AP:IgG Molar Ratio

This kit utilizes a 2:1 AP:IgG molar ratio which is optimal for most conjugations reaction. However, lower or higher ratios may give better results depending upon the antibody characteristics and the intended end-use. To change the AP:IgG molar ratio, vary the volume of AP-ZTM added to the conjugation reaction (Step 8).

CONJUGATION PROCEDURE - OVERVIEW





BEFORE BEGINNING THE PROCEDURE

Remove the Concentrated Activator from the freezer. <u>Important:</u> Allow sufficient time to let the <u>container and contents come to room temperature before opening the outer and inner vials.</u>

Note: The jar containing the Activator can be removed from the freezer up to 24 hours before use.

DETAILED CONJUGATION PROCEDURE

- 1. Measure the absorbance of the IgG solution at 280 nm using PBS as a blank. Divide the A280 by 1.40 to obtain the IgG concentration in mg/ml.
- 2. Dilute IgG to 1.20 mg/ml in 1X PBS (0.80 1.4 mg/ml is acceptable).
- 3. Add 10 mg of IgG solution to a new microcentrifuge tube.
- 4. Prepare a working dilution (1X) of Activator from Concentrated Activator in deionized water:
 - a. Add 4.0 µL of Concentrated Activator to 120 µL of deionized water.
 - b. Immediately vortex to mix the solution thoroughly.

Note: The <u>1X</u> Activator must be used within 5 minutes of preparation. If more than 5 minutes passes before use, discard the 1X Activator and prepare a fresh solution.

- 5. Add 100 μ L of <u>1X</u> Activator to the 10 mg aliquot of IgG and then mix thoroughly by gentle vortexing.
- 6. Incubate the solution at room temperature for 1 hour.

Note: A longer incubation is not harmful and even overnight incubations will be successful.

7. Desalt the complete reaction volume into pure 1X PBS. We recommend using Pierce Zeba desalting spin columns with a 7 Kd MW cutoff for small volumes of IgG. Use of gravity desalting columns, dialysis, and extensive washing with centrifugal filter units for desalting is also acceptable.

Note: The activated IgG is stable and can be stored at 2-8°C for at least 4 months.

- 8. Add 2 mL of AP-ZTM to the desalted, activated IgG and mix by gentle vortexing.
- 9. Incubate the solution at room temperature for 2-24 hours.

Note: Usable conjugates are produced after only 2 hours incubation. Larger and more potent conjugates will be produced after longer incubations.

- 10. Add 400 µL of Quenching Reagent to the reaction and mix by gentle vortexing.
- 11. Incubate the solution at room temperature for 1 hour.

Note: A longer incubation is not harmful and overnight incubations will be successful.

12. Conjugate is ready for use. Store at 2-8°C.

Note: To improve conjugate performance, it may help to purify the conjugate from the unincorporated AP and reaction components by size exclusion chromatography.



RECOMMENDED ACCESSORIES

For desalting IgG after activation - Order from ThermoFisher:

| Sample Size | Description | Cat # |
|-----------------|--|--------------|
| 2 – 12 μL | Zeba Spin Desalting Columns, Micro (75µL), 7K MWCO | 89877, 89878 |
| 30 - 130 μL | Zeba Spin Desalting Columns, 0.5mL, 7K MWCO | 89882, 89883 |
| $200-700~\mu L$ | Zeba Spin Desalting Columns, 2mL, 7K MWCO | 89889, 89890 |
| 500 – 2000 μL | Zeba Spin Desalting Columns, 5mL, 7K MWCO | 89891, 89892 |
| 700 – 4000 μL | Zeba Spin Desalting Columns, 10mL, 7K MWCO | 89893, 89894 |

<u>For concentrating IgG before or after activation or for concentrating the final conjugate – Order from MilliporeSigma:</u>

| Sample Size | Description | Cat# |
|--------------|--|-----------|
| Up to 500 μL | Amicon Ultra-0.5 Centrifugal Filter Unit with Ultracel-50 membrane | Z740176 |
| Up to 2 mL | Amicon Ultra-2 Centrifugal Filter Unit with Ultracel-50 membrane | UFC205024 |
| Up to 4 mL | Amicon Ultra-4 Centrifugal Filter Unit with Ultracel-50 membrane | UFC805008 |
| Up to 15 mL | Amicon Ultra-15 Centrifugal Filter Unit with Ultracel-50 membrane | Z648000 |